

# South Gas Company

Email: [south\\_gascom@sgciraq.com](mailto:south_gascom@sgciraq.com)

Website: [www.sgciraq.com](http://www.sgciraq.com)

## SGC-M28-2011

ITEM	DESCRIPTION	UNIT	QTY
	<b>TEST EQUIPMENT USED FOR DEHIDRATION &amp; COMPRESS STATIONS AT SOUTH , NORTH RUMAILA AND ZUBAIR</b>		
1	PASCAL 100 PORTABLE MULTI FUNCTION CALIBRATOR FOR TESTING OF PNEUMATIC , ELECTRONIC & TEMP. PROCESS INSTRUMENTS. - MODEL : P100-IX-0X-P2-X <b>Or Equivalent</b> - ONE INPUT MODULE &ONE OUTPUT MODULE. - 2 INTERNAL RANGES : 212G & 500G - WITH ONE EXTERNAL TRANSDUCER PSP-1/100	NOS	6
2	PASCAL ET PORTABLE MULTI FUNCTION CALIBRATOR FOR TESTING OF ELECTRONIC & TEMP. PROCESS INSTRUMENTS. - MODEL : ET-IX-0X-X-X <b>Or Equivalent</b> - ONE INPUT MODULE &ONE OUTPUT MODULE.	NOS	6
3	HYDROLIC PRESSURE GENERATOR - MODEL : GHM 700 <b>Or Equivalent</b> WITH ACCESSORIES : - NF4 - NF2 - BF2 - NM4 - NM2	NOS	6
4	PRESSURE & VACUUM PNEUMATIC GENERATOR - MODEL : GPP40 <b>Or Equivalent</b> - RANGE : - 0.9 TO 40 BAR WITH ACCESSORIES : - NF4 - NF2 - BF2 - NM4 - NM2	NOS	6
5	PORTABLE DRY BLOCK FOR TEMP. CALIBRATION - MODEL : BL - 7 <b>Or Equivalent</b> - RANGE : AMBIENT + 700 °C	NOS	6

6	<p>HYDROLIC COMPARATOR FOR PRESSURE CALIBRATION</p> <ul style="list-style-type: none"> <li>- MODEL : MPA-C <b>Or Equivalent</b></li> <li>- RANGE : UP TO 1200 BAR</li> <li>- WITH / 0.5L HYDROLIC OIL</li> <li>- WITH / ONE SET OF SPANNERS 30/32</li> <li>- WITH /ONE SET OF ROTATING CONNECTORS FOR GAUGES</li> <li>- WITH/ ONE SET OF SEALS</li> <li>- WITH / TWO THREADED LOCKING RING</li> <li>- WITH / ONE SET OF BSP ADAPTER KIT CODE NO.: 241086</li> <li>- WITH / ONE SET OF NPT ADAPTER KIT CODE NO.: 241083</li> <li>- WITH / ONE SET OF LOW PRESSURE ADAPTER KIT CODE NO.: 241084</li> <li>- WITH/ ONE SET OF METRIC ADAPTER KIT CODE NO.: 241085</li> </ul>	NOS	12
7	<p>HAND HELD T/C &amp; RTD CALIBRATOR</p> <ul style="list-style-type: none"> <li>- MODEL : G40 / TR <b>Or Equivalent</b></li> </ul> <p>WITH ACCESSORIES :</p> <ul style="list-style-type: none"> <li>- SOFT BAG WITH SHOLDER STRAP BM-001</li> <li>- SET OF THREE PLUGE FOR THERMO COUPLES TYPE J / K / S</li> </ul>	NOS	10
8	<p>HAND HELD CURRENT LOOP (4 - 20 ) ma CALIBRATOR</p> <ul style="list-style-type: none"> <li>- MODEL : G40 / AM <b>Or Equivalent</b></li> </ul> <p>WITH ACCESSORIES :</p> <ul style="list-style-type: none"> <li>- SOFT BAG WITH SHOLDER STRAP BM-001</li> </ul>	NOS	15
9	<p>SET OF PRECISION TEST GAUGES 1150C SERIES <b>Or Equivalent</b></p> <p>DIAL SIZE : 6"</p> <p>ACCURACY: 0.25% f.s</p> <p>MOUNTING: LOWER CONNECTION</p> <p>WITH CARRYING CASE.</p> <p>RANGES:</p> <ul style="list-style-type: none"> <li>- 0 – 2.5 BAR</li> <li>- 0 – 16 BAR</li> <li>- 0 – 100 BAR</li> </ul>	SET	6
10	<p>CALIBRATOR FOR PNEUMATIC INSTRUMENT</p> <p>MODEL: VE-T-PG-5 <b>Or Equivalent</b></p> <ul style="list-style-type: none"> <li>- RANGE : 0 – 20 PSI</li> <li>- UNCERTAINTY :0.5 % f.s</li> <li>- WITH TWO GAUGES</li> <li>- BUILT IN HAND PUMP ( PRESSURE GENERATOR) FOR MAX PRESS 7 BAR AND VACCUM - 800mbar</li> <li>- LIDS RANGE: 0 – 20 PSI UNCERTAINTY 0.5 % f.s WITH TWO GAUGES.</li> </ul>	NOS	3
11	<ul style="list-style-type: none"> <li>- TRAINING COURSE FOR USING THE ABOVE SCANDORA TEST EQUIPMENTS FOR 7 DAYS / 4 TECHNICIAN IN VENDOR WORKSHOP</li> </ul>	LOT	1
	<p><b>NOTE:</b></p> <p><b>1-THE OFFERS SHOULD CONTAINS COMPLETE DATA SHEET, SPESIFICATION &amp; DRAWING FROM THE MANUFACTURER.</b></p> <p><b>2-THE VENDORS SHOULD SUPPLY CERTIFICATION AND SEIRAL NUMBER FOR MANUFACTURE FOR EACH ITEM.</b></p> <p><b>3- WARRANTY FOR ONE YEAR</b></p>		
	<p><b>County of Origin: USA , West Europe, JAPAN</b></p>		